



## E006: Temperature monitoring devices

<b>PQS code:</b>	<b>E006/100</b>
<b>Description:</b>	User-programmable temperature data logger
<b>Manufacturer's ref:</b>	LMPPro T051
<b>Manufactured in:</b>	India
<b>Company:</b>	G-Tek Corporation Private Limited
<b>Address:</b>	3, MAHAVIR INDUSTRIAL ESTATE NEAR JALARAM MANDIR, Karelibagh, Vadodara - 390018 Gujarat, India
<b>Telephone:</b>	+91 98 242 61702
<b>Email:</b>	<a href="mailto:asu@gtek-india.com">asu@gtek-india.com</a>
<b>Web address:</b>	<a href="http://www.gtek-india.com">www.gtek-india.com</a>

## Specifications

<b>Product type:</b>	User-programmable temperature data logger	<b>Display medium:</b>	LCD display
<b>Mode of operation:</b>	Electronic	<b>Weight</b>	0.051
<b>Minimum temperature:</b>	-30	<b>Main material:</b>	Polycarbonate plastic; non-breakable, non-corrodible
<b>Maximum temperature:</b>	+60 °C	<b>Dimensions:</b>	9.55 (L) x 4.5 (W) x 1.48 (H) cm

### MECHANICAL AND ELECTRONIC DEVICES ONLY

<b>Accuracy:</b>	± 0.5 °C for -20 to +30 °C ± 0.7 °C Otherwise	<b>Min logging interval:</b>	5 second
<b>Power source:</b>	Lithium Coin Cell Battery	<b>Alarm type:</b>	Visual
<b>Battery type:</b>	User Replaceable 3.0 V 225 mAH; CR2032 Panasonic (	<b>User interface:</b>	Keyboard Interface
<b>Battery shelf life:</b>	1 year	<b>Programmability:</b>	User programmed
<b>Activated life:</b>	12 months	<b>IP rating:</b>	IP 54
<b>Memory capacity:</b>	50000		

<b>Comments:</b>	LMPPro T051 User Programmable Temperature Data logger is an Internal sensor data logger meeting the requirements of WHO PQS E006/TR05.1. It can store up to 50000 data records. User can fully configure parameters like – Start delay, alarm high/low with corresponding alarm delay, alarm event type- single/cumulative, temperature unit, UTC time zone, store interval, Batch type- Stop on full or memory rollover etc. User can also set the information like, username and job description. It displays OK symbol "Tick" if the temperature has remained within the recommended limits and NOK symbol "Cross" if there are violations for the set alarm limits. User can also download data from multiple data loggers in single file and generate report in pdf format or export data as a csv file.
------------------	---

<b>Minimum order:</b>	1	<b>Year base price:</b>	2024
<b>Quality standard:</b>	- ISO 9001:2015 - Other -	<b>Pieces per carton:</b>	25
<b>Price per unit:</b>	For 1-10 units US\$49 per unit For 100-250 units US\$34 per unit For more than 250 units :Consult Factory	<b>Volume per carton:</b>	0.018
		<b>Weight per carton:</b>	2.5
		<b>Incoterms:</b>	EXW
<b>Verification report:</b>	LLTPL/TL/2024-02/01	<b>Verification laboratory:</b>	Lisaline Lifescience Technologies PVT. LTD
<b>Current PQS status:</b>	pre-qualified ;	<b>Valid until:</b>	May 2026

Note: If Current PQS status is 'Suspended' or 'Withdrawn', this product is NOT to be purchased.